



January 30, 2024

William Kotas
Intertek PSI
17 British American Boulevard
Latham, NY 12110

RE: Project: MECHANICVILLE CENTRAL SCHOOL D
Pace Project No.: 70285337

Dear William Kotas:

Enclosed are the analytical results for sample(s) received by the laboratory on January 26, 2024. The results relate only to the samples included in this report. Results reported herein conform to the applicable TNI/NELAC Standards and the laboratory's Quality Manual, where applicable, unless otherwise noted in the body of the report.

The test results provided in this final report were generated by each of the following laboratories within the Pace Network:

- Pace Analytical Services - Melville

If you have any questions concerning this report, please feel free to contact me.

Sincerely,

Lori A. Beyer
lori.beyer@pacelabs.com
516-370-6014
Project Manager

Enclosures



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



CERTIFICATIONS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Pace Analytical Services Long Island

575 Broad Hollow Rd, Melville, NY 11747

Connecticut Certification #: PH-0435

Delaware Certification # NY 10478

Maryland Certification #: 208

Massachusetts Certification #: M-NY026

New Hampshire Certification #: 2987

New Jersey Certification #: NY158

New York Certification #: 10478 Primary Accrediting Body

Pennsylvania Certification #: 68-00350

Rhode Island Certification #: LAO00340

Virginia Certification # 460302

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



SAMPLE SUMMARY

Project: MECHANICVILLE CENTRAL SCHOOL D
Pace Project No.: 70285337

Lab ID	Sample ID	Matrix	Date Collected	Date Received
70285337001	MHS-KK-S	Drinking Water	01/25/24 06:01	01/26/24 08:00
70285337002	MHS-KK-1M	Drinking Water	01/25/24 06:01	01/26/24 08:00
70285337003	MHS-KP-S1	Drinking Water	01/25/24 06:02	01/26/24 08:00
70285337004	MHS-KP-S2	Drinking Water	01/25/24 06:03	01/26/24 08:00
70285337005	MHS-K-S	Drinking Water	01/25/24 06:06	01/26/24 08:00
70285337006	MHS-170	Drinking Water	01/25/24 06:13	01/26/24 08:00
70285337007	MHS-169 (FTN)	Drinking Water	01/25/24 06:13	01/26/24 08:00
70285337008	MHS-169A (B.F)	Drinking Water	01/25/24 06:13	01/26/24 08:00
70285337009	MHS-143	Drinking Water	01/25/24 06:16	01/26/24 08:00
70285337010	MHS-144 (FTN)	Drinking Water	01/25/24 06:18	01/26/24 08:00
70285337011	MHS-144A (B.F)	Drinking Water	01/25/24 06:18	01/26/24 08:00
70285337012	MHS-145	Drinking Water	01/25/24 06:18	01/26/24 08:00
70285337013	MHS-68	Drinking Water	01/25/24 06:29	01/26/24 08:00
70285337014	MHS-72 (FTN)	Drinking Water	01/25/24 06:27	01/26/24 08:00
70285337015	MHS-72A (B.F.)	Drinking Water	01/25/24 06:27	01/26/24 08:00
70285337016	MHS-69	Drinking Water	01/25/24 06:28	01/26/24 08:00
70285337017	MHS-73 (FTN)	Drinking Water	01/25/24 06:31	01/26/24 08:00
70285337018	MHS-73A (B.F.)	Drinking Water	01/25/24 06:31	01/26/24 08:00
70285337019	MHS-95	Drinking Water	01/25/24 06:42	01/26/24 08:00
70285337020	MHS-96	Drinking Water	01/25/24 06:42	01/26/24 08:00
70285337021	MHS-97	Drinking Water	01/25/24 06:42	01/26/24 08:00
70285337022	MHS-98	Drinking Water	01/25/24 06:42	01/26/24 08:00
70285337023	MHS-99	Drinking Water	01/25/24 06:42	01/26/24 08:00
70285337024	MHS-111	Drinking Water	01/25/24 06:50	01/26/24 08:00
70285337025	MHS-111A (B.F.)	Drinking Water	01/25/24 06:50	01/26/24 08:00
70285337026	MHS-112	Drinking Water	01/25/24 06:55	01/26/24 08:00
70285337027	MHS-117	Drinking Water	01/25/24 06:56	01/26/24 08:00
70285337028	MHS-300C-F1	Drinking Water	01/25/24 07:01	01/26/24 08:00
70285337029	MHS-300C-BF1	Drinking Water	01/25/24 07:01	01/26/24 08:00
70285337030	MHS-300C-F2	Drinking Water	01/25/24 07:01	01/26/24 08:00
70285337031	MHS-301C-F1	Drinking Water	01/25/24 07:06	01/26/24 08:00
70285337032	MHS-301C-BF1	Drinking Water	01/25/24 07:06	01/26/24 08:00
70285337033	MHS-301C-F2	Drinking Water	01/25/24 07:06	01/26/24 08:00
70285337034	MHS-301C-BF2	Drinking Water	01/25/24 07:06	01/26/24 08:00
70285337035	MHS-302-1	Drinking Water	01/25/24 07:15	01/26/24 08:00
70285337036	MHS-302-2	Drinking Water	01/25/24 07:15	01/26/24 08:00
70285337037	MHS-302-3	Drinking Water	01/25/24 07:15	01/26/24 08:00

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



SAMPLE SUMMARY

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Lab ID	Sample ID	Matrix	Date Collected	Date Received
70285337038	MHS-WRF	Drinking Water	01/25/24 07:20	01/26/24 08:00
70285337039	MHS-162	Drinking Water	01/25/24 07:21	01/26/24 08:00
70285337040	MHS-163	Drinking Water	01/25/24 07:21	01/26/24 08:00
70285337041	MHS-2 (FTN)	Drinking Water	01/25/24 07:25	01/26/24 08:00
70285337042	MHS-2 A(B.F.)	Drinking Water	01/25/24 07:25	01/26/24 08:00
70285337043	BG-S2	Drinking Water	01/25/24 07:41	01/26/24 08:00
70285337044	BG-1	Drinking Water	01/25/24 07:42	01/26/24 08:00

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

**SAMPLE ANALYTE COUNT**

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Lab ID	Sample ID	Method	Analysts	Analytes Reported
70285337001	MHS-KK-S	EPA 200.8	JJS	1
70285337002	MHS-KK-1M	EPA 200.8	JJS	1
70285337003	MHS-KP-S1	EPA 200.8	JJS	1
70285337004	MHS-KP-S2	EPA 200.8	JJS	1
70285337005	MHS-K-S	EPA 200.8	JJS	1
70285337006	MHS-170	EPA 200.8	JJS	1
70285337007	MHS-169 (FTN)	EPA 200.8	JJS	1
70285337008	MHS-169A (B.F)	EPA 200.8	JJS	1
70285337009	MHS-143	EPA 200.8	JJS	1
70285337010	MHS-144 (FTN)	EPA 200.8	JJS	1
70285337011	MHS-144A (B.F)	EPA 200.8	JJS	1
70285337012	MHS-145	EPA 200.8	JJS	1
70285337013	MHS-68	EPA 200.8	JJS	1
70285337014	MHS-72 (FTN)	EPA 200.8	JJS	1
70285337015	MHS-72A (B.F.)	EPA 200.8	JJS	1
70285337016	MHS-69	EPA 200.8	JJS	1
70285337017	MHS-73 (FTN)	EPA 200.8	JJS	1
70285337018	MHS-73A (B.F.)	EPA 200.8	JJS	1
70285337019	MHS-95	EPA 200.8	JJS	1
70285337020	MHS-96	EPA 200.8	JJS	1
70285337021	MHS-97	EPA 200.8	JJS	1
70285337022	MHS-98	EPA 200.8	JJS	1
70285337023	MHS-99	EPA 200.8	JJS	1
70285337024	MHS-111	EPA 200.8	JJS	1
70285337025	MHS-111A (B.F.)	EPA 200.8	JJS	1
70285337026	MHS-112	EPA 200.8	JJS	1
70285337027	MHS-117	EPA 200.8	JJS	1
70285337028	MHS-300C-F1	EPA 200.8	JJS	1
70285337029	MHS-300C-BF1	EPA 200.8	JJS	1
70285337030	MHS-300C-F2	EPA 200.8	JJS	1
70285337031	MHS-301C-F1	EPA 200.8	JJS	1
70285337032	MHS-301C-BF1	EPA 200.8	JJS	1
70285337033	MHS-301C-F2	EPA 200.8	JJS	1
70285337034	MHS-301C-BF2	EPA 200.8	JJS	1
70285337035	MHS-302-1	EPA 200.8	JJS	1
70285337036	MHS-302-2	EPA 200.8	JJS	1
70285337037	MHS-302-3	EPA 200.8	JJS	1

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



SAMPLE ANALYTE COUNT

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Lab ID	Sample ID	Method	Analysts	Analytes Reported
70285337038	MHS-WRF	EPA 200.8	JJS	1
70285337039	MHS-162	EPA 200.8	JJS	1
70285337040	MHS-163	EPA 200.8	JJS	1
70285337041	MHS-2 (FTN)	EPA 200.8	JJS	1
70285337042	MHS-2 A(B.F.)	EPA 200.8	JJS	1
70285337043	BG-S2	EPA 200.8	JJS	1
70285337044	BG-1	EPA 200.8	JJS	1

PACE-MV = Pace Analytical Services - Melville

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-KK-S								
Lab ID: 70285337001								
Collected: 01/25/24 06:01								
Received: 01/26/24 08:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		01/29/24 14:38	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-KK-1M	Lab ID: 70285337002	Collected: 01/25/24 06:01	Received: 01/26/24 08:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		01/29/24 14:42	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-KP-S1	Lab ID: 70285337003	Collected: 01/25/24 06:02	Received: 01/26/24 08:00	Matrix: Drinking Water					
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		01/29/24 14:44	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-KP-S2		Lab ID: 70285337004	Collected: 01/25/24 06:03	Received: 01/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		01/29/24 14:45	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-K-S		Lab ID: 70285337005	Collected: 01/25/24 06:06	Received: 01/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	1.7	ug/L	1.0	1		01/29/24 15:36	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-170								
Lab ID: 70285337006								
Collected: 01/25/24 06:13								
Received: 01/26/24 08:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		01/29/24 15:40	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-169 (FTN)								
Lab ID: 70285337007								
Collected: 01/25/24 06:13 Received: 01/26/24 08:00 Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		01/29/24 15:45	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-169A (B.F)								
Lab ID: 70285337008								
Collected: 01/25/24 06:13 Received: 01/26/24 08:00 Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		01/29/24 15:46	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-143								
Lab ID: 70285337009								
Collected: 01/25/24 06:16								
Received: 01/26/24 08:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.1	ug/L	1.0	1		01/29/24 15:48	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-144 (FTN)								
Lab ID: 70285337010								
Collected: 01/25/24 06:18								
Received: 01/26/24 08:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		01/29/24 15:49	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-144A (B.F)								
Lab ID: 70285337011								
Collected: 01/25/24 06:18								
Received: 01/26/24 08:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		01/29/24 15:54	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-145	Lab ID: 70285337012	Collected: 01/25/24 06:18	Received: 01/26/24 08:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		01/29/24 15:55	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-68								
Lab ID: 70285337013								
Collected: 01/25/24 06:29 Received: 01/26/24 08:00 Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	6.6	ug/L	1.0	1		01/29/24 15:57	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-72 (FTN)		Lab ID: 70285337014	Collected: 01/25/24 06:27	Received: 01/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		01/29/24 15:59	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-72A (B.F.)								
Lab ID: 70285337015								
Collected: 01/25/24 06:27								
Received: 01/26/24 08:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		01/29/24 16:00	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-69								
Lab ID: 70285337016								
Collected: 01/25/24 06:28								
Received: 01/26/24 08:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	1.9	ug/L	1.0	1		01/29/24 16:02	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-73 (FTN)		Lab ID: 70285337017	Collected: 01/25/24 06:31	Received: 01/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		01/29/24 16:03	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-73A (B.F.)	Lab ID: 70285337018	Collected: 01/25/24 06:31	Received: 01/26/24 08:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		01/29/24 16:05	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-95								
Lab ID: 70285337019								
Collected: 01/25/24 06:42								
Received: 01/26/24 08:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	1.8	ug/L	1.0	1		01/29/24 16:06	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-96								
Lab ID: 70285337020								
Collected: 01/25/24 06:42 Received: 01/26/24 08:00 Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.9	ug/L	1.0	1		01/29/24 16:08	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-97								
Lab ID: 70285337021								
Collected: 01/25/24 06:42 Received: 01/26/24 08:00 Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	2.7	ug/L	1.0	1		01/29/24 16:12	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-98								
Lab ID: 70285337022								
Collected: 01/25/24 06:42 Received: 01/26/24 08:00 Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	6.1	ug/L	1.0	1		01/29/24 16:14	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-99								
Lab ID: 70285337023								
Collected: 01/25/24 06:42 Received: 01/26/24 08:00 Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	2.4	ug/L	1.0	1		01/29/24 16:15	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-111	Lab ID: 70285337024	Collected: 01/25/24 06:50	Received: 01/26/24 08:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		01/29/24 16:17	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-111A (B.F.)	Lab ID: 70285337025	Collected: 01/25/24 06:50	Received: 01/26/24 08:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		01/29/24 16:21	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-112		Lab ID: 70285337026	Collected: 01/25/24 06:55	Received: 01/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		01/29/24 16:26	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-117	Lab ID: 70285337027	Collected: 01/25/24 06:56	Received: 01/26/24 08:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		01/29/24 16:33	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-300C-F1								
Lab ID: 70285337028								
Collected: 01/25/24 07:01 Received: 01/26/24 08:00 Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	1.1	ug/L	1.0	1		01/29/24 16:35	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-300C-BF1								
Lab ID: 70285337029								
Collected: 01/25/24 07:01								
Received: 01/26/24 08:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	1.1	ug/L	1.0	1		01/29/24 16:36	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-300C-F2	Lab ID: 70285337030	Collected: 01/25/24 07:01	Received: 01/26/24 08:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.2	ug/L	1.0	1		01/29/24 16:38	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-301C-F1	Lab ID: 70285337031	Collected: 01/25/24 07:06	Received: 01/26/24 08:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		01/29/24 16:39	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-301C-BF1								
Lab ID: 70285337032								
Collected: 01/25/24 07:06								
Received: 01/26/24 08:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		01/29/24 16:41	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-301C-F2								
Lab ID: 70285337033								
Collected: 01/25/24 07:06								
Received: 01/26/24 08:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		01/29/24 16:43	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-301C-BF2								
Lab ID: 70285337034								
Collected: 01/25/24 07:06								
Received: 01/26/24 08:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		01/29/24 16:44	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-302-1	Lab ID: 70285337035	Collected: 01/25/24 07:15	Received: 01/26/24 08:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	48.7	ug/L	1.0	1		01/29/24 16:49	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-302-2								
Lab ID: 70285337036								
Collected: 01/25/24 07:15 Received: 01/26/24 08:00 Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	60.8	ug/L	1.0	1		01/29/24 16:50	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-302-3								
Lab ID: 70285337037								
Collected: 01/25/24 07:15								
Received: 01/26/24 08:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	8.4	ug/L	1.0	1		01/29/24 16:52	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-WRF		Lab ID: 70285337038	Collected: 01/25/24 07:20	Received: 01/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		01/29/24 16:53	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-162	Lab ID: 70285337039	Collected: 01/25/24 07:21	Received: 01/26/24 08:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		01/29/24 16:55	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: MHS-163								
Lab ID: 70285337040								
Collected: 01/25/24 07:21								
Received: 01/26/24 08:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	<1.0	ug/L	1.0	1		01/29/24 16:56	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-2 (FTN)	Lab ID: 70285337041	Collected: 01/25/24 07:25	Received: 01/26/24 08:00	Matrix: Drinking Water				
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		01/29/24 16:58	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: MHS-2 A(B.F.)		Lab ID: 70285337042	Collected: 01/25/24 07:25	Received: 01/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		01/29/24 16:59	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: BG-S2		Lab ID: 70285337043	Collected: 01/25/24 07:41	Received: 01/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	9.4	ug/L	1.0	1		01/29/24 17:01	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Sample: BG-1		Lab ID: 70285337044	Collected: 01/25/24 07:42	Received: 01/26/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		01/29/24 17:02	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

QC Batch:	335549	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville

Associated Lab Samples: 70285337001, 70285337002, 70285337003, 70285337004

METHOD BLANK: 1724107 Matrix: Water
 Associated Lab Samples: 70285337001, 70285337002, 70285337003, 70285337004

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	01/29/24 14:06	

LABORATORY CONTROL SAMPLE: 1724108

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	49.8	100	85-115	

MATRIX SPIKE SAMPLE: 1724110

Parameter	Units	70285276001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	47.3	94	70-130	

MATRIX SPIKE SAMPLE: 1724112

Parameter	Units	70285438001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	48.1	96	70-130	

SAMPLE DUPLICATE: 1724109

Parameter	Units	70285276001 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		20	

SAMPLE DUPLICATE: 1724111

Parameter	Units	70285438001 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		20	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

QC Batch:	335550	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70285337005, 70285337006, 70285337007, 70285337008, 70285337009, 70285337010, 70285337011, 70285337012, 70285337013, 70285337014, 70285337015, 70285337016, 70285337017, 70285337018, 70285337019, 70285337020, 70285337021, 70285337022, 70285337023, 70285337024		

METHOD BLANK:	1724114	Matrix:	Water
Associated Lab Samples:	70285337005, 70285337006, 70285337007, 70285337008, 70285337009, 70285337010, 70285337011, 70285337012, 70285337013, 70285337014, 70285337015, 70285337016, 70285337017, 70285337018, 70285337019, 70285337020, 70285337021, 70285337022, 70285337023, 70285337024		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	01/29/24 15:30	

LABORATORY CONTROL SAMPLE:	1724115					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	47.7	95	85-115	

MATRIX SPIKE SAMPLE:	1724117						
Parameter	Units	70285337005 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	1.7	50	49.2	95	70-130	

MATRIX SPIKE SAMPLE:	1724119						
Parameter	Units	70285337006 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	44.6	89	70-130	

SAMPLE DUPLICATE:	1724116					
Parameter	Units	70285337005 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	1.7	1.7	0	20	

SAMPLE DUPLICATE:	1724118					
Parameter	Units	70285337006 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		20	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

QC Batch:	335551	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70285337025, 70285337026, 70285337027, 70285337028, 70285337029, 70285337030, 70285337031, 70285337032, 70285337033, 70285337034, 70285337035, 70285337036, 70285337037, 70285337038, 70285337039, 70285337040, 70285337041, 70285337042, 70285337043, 70285337044		

METHOD BLANK:	1724121	Matrix:	Water
Associated Lab Samples:	70285337025, 70285337026, 70285337027, 70285337028, 70285337029, 70285337030, 70285337031, 70285337032, 70285337033, 70285337034, 70285337035, 70285337036, 70285337037, 70285337038, 70285337039, 70285337040, 70285337041, 70285337042, 70285337043, 70285337044		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	01/29/24 16:18	

LABORATORY CONTROL SAMPLE:	1724122					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	49.9	100	85-115	

MATRIX SPIKE SAMPLE:	1724124						
Parameter	Units	70285337025 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	44.3	88	70-130	

MATRIX SPIKE SAMPLE:	1724126						
Parameter	Units	70285337026 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	45.2	89	70-130	

SAMPLE DUPLICATE:	1724123					
Parameter	Units	70285337025 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		20	

SAMPLE DUPLICATE:	1724125					
Parameter	Units	70285337026 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		20	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALIFIERS

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Reported results are not rounded until the final step prior to reporting. Therefore, calculated parameters that are typically reported as "Total" may vary slightly from the sum of the reported component parameters.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: MECHANICVILLE CENTRAL SCHOOL D

Pace Project No.: 70285337

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70285337001	MHS-KK-S	EPA 200.8	335549		
70285337002	MHS-KK-1M	EPA 200.8	335549		
70285337003	MHS-KP-S1	EPA 200.8	335549		
70285337004	MHS-KP-S2	EPA 200.8	335549		
70285337005	MHS-K-S	EPA 200.8	335550		
70285337006	MHS-170	EPA 200.8	335550		
70285337007	MHS-169 (FTN)	EPA 200.8	335550		
70285337008	MHS-169A (B.F)	EPA 200.8	335550		
70285337009	MHS-143	EPA 200.8	335550		
70285337010	MHS-144 (FTN)	EPA 200.8	335550		
70285337011	MHS-144A (B.F)	EPA 200.8	335550		
70285337012	MHS-145	EPA 200.8	335550		
70285337013	MHS-68	EPA 200.8	335550		
70285337014	MHS-72 (FTN)	EPA 200.8	335550		
70285337015	MHS-72A (B.F.)	EPA 200.8	335550		
70285337016	MHS-69	EPA 200.8	335550		
70285337017	MHS-73 (FTN)	EPA 200.8	335550		
70285337018	MHS-73A (B.F.)	EPA 200.8	335550		
70285337019	MHS-95	EPA 200.8	335550		
70285337020	MHS-96	EPA 200.8	335550		
70285337021	MHS-97	EPA 200.8	335550		
70285337022	MHS-98	EPA 200.8	335550		
70285337023	MHS-99	EPA 200.8	335550		
70285337024	MHS-111	EPA 200.8	335550		
70285337025	MHS-111A (B.F.)	EPA 200.8	335551		
70285337026	MHS-112	EPA 200.8	335551		
70285337027	MHS-117	EPA 200.8	335551		
70285337028	MHS-300C-F1	EPA 200.8	335551		
70285337029	MHS-300C-BF1	EPA 200.8	335551		
70285337030	MHS-300C-F2	EPA 200.8	335551		
70285337031	MHS-301C-F1	EPA 200.8	335551		
70285337032	MHS-301C-BF1	EPA 200.8	335551		
70285337033	MHS-301C-F2	EPA 200.8	335551		
70285337034	MHS-301C-BF2	EPA 200.8	335551		
70285337035	MHS-302-1	EPA 200.8	335551		
70285337036	MHS-302-2	EPA 200.8	335551		
70285337037	MHS-302-3	EPA 200.8	335551		
70285337038	MHS-WRF	EPA 200.8	335551		
70285337039	MHS-162	EPA 200.8	335551		
70285337040	MHS-163	EPA 200.8	335551		
70285337041	MHS-2 (FTN)	EPA 200.8	335551		
70285337042	MHS-2 A(B.F.)	EPA 200.8	335551		
70285337043	BG-S2	EPA 200.8	335551		
70285337044	BG-1	EPA 200.8	335551		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.

CHAIN-OF-CUSTODY Analytical Request Document
 Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

LAB USE ONLY - Affix Workorder/Label Here
WO#: 70285337

70285337

Company Name: **Intertek-PSI**
 Street Address: **17 British American Blvd, Latham NY 12110**

Contact/Report To: **William A. Kotas**
 Phone #: **518-377-9841**
 E-Mail: William.Kotas@intertek.com
 Cc E-Mail:

Invoice To: **PSI Latham NY Accounts Payable**
 Invoice E-Mail: LathamAR@intertek.com

Purchase Order # (if applicable): **CR Boces**
 Quote #: **CR Boces**

Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET
 Date Deliverables: [] Level II [] Level III [] Level IV
 [] EQUIS [] Other
 Date Results Requested: **Standard 10 business day**

Regulatory Program (DW, RCRA, etc.) as applicable: **NY Lead in School DW**
 Rush (Pre-approval required): **DW PWSID # or WW Permit # as applicable:**
 [] 2 Day [] 3 day [] 5 day [] Other
 Field Filtered (if applicable): [] Yes [] No
 Analysis: **Standard 10 business day**

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start)		Composite End		Res. CL2	Number & Type of Containers		Preservation non-conformance identified for sample.
			Date	Time	Date	Time		Plastic	Glass	
MHS - KK - S	DW	G	1/25/24	6:01				1		
MHS - K - IM				6:01						
MHS - KP - S1				6:02						
MHS - KP - S2				6:03						
MHS - K - S				6:06						
MHS - 170				6:13						
MHS 164 (FTN)				6:13						
MHS 169A (B.F.)				6:13						
MHS 143				6:16						
MHS 144 (FTN)				6:18						

Customer Remarks / Special Conditions / Possible Hazards:

Collected By: **William A. Kotas (PSI)**
 Signature: *[Signature]*

Coolers: **3** Thermometer ID: **74711** Correction Factor (°C): **20.2** Obs. Temp (°C): **20.2** Corrected Temp (°C): **20.6**

Tracking Number: **17214**

Date/Time: **1/25/24 12:01** Received by/Company: **PSI**
 Date/Time: **1/25/24 12:14** Received by/Company: **MPIN**
 Date/Time: **1/26 4:10** Received by/Company: **MPIN**
 Date/Time: **1/26 8:00** Received by/Company: **MPIN**

Delivered by: [] In-Person [] Courier [] FedEx [] UPS [] Other

Page: **1** of **5**

CHAIN-OF-CUSTODY Analytical Request Document
 Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields



Scan QR Code for Instructions

Pace* Location Requested (City/State): Pace Analytical Long Island NY, 575 Broad Hollow Rd, Melville, NY 11747

Company Name: Intertek-PSI
Street Address: 17 British American Blvd, Latham NY 12110

Contact/Report To: William A. Kotas
Phone #: 518-377-9841
E-Mail: William.Kotas@intertek.com
Cc E-Mail:

Invoice To: PSI Latham NY Accounts Payable
Invoice E-Mail: LathamAR@intertek.com

Purchase Order # (if applicable): CR Bores
Quote #: CR Bores

County / State origin of sample(s): New York

Regulatory Program (DW, RCHA, etc.) as applicable: NY Lead in School DW

Rush (Pre-approval required): DW PWSID # or WW Permit # as applicable:
 12 Day 3 day 15 day Other _____
Date Results Requested: Standard 10 business day

Field Filtered (if applicable): Yes No

Analysis:

Time Zone Collected: JAK JPT JMT JCT JET

Data Deliverables:

Level II Level III Level IV

EQUIS

Other _____

*** Matrix Codes (Insert in Matrix box below):** Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start)		Res. CLZ	Number & Type of Containers	Sample Comment
			Date	Time			
MHS 147A (B.F.)	DW	G	1/25/24	6:18		1	200.8 Drinking Water (Pb only)
MHS 145				6:18			
MHS 68				6:29			
MHS 72 (FTN.)				6:27			
MHS 72A (B.F.)				6:27			
MHS 69				6:28			
MHS 73 (FTN.)				6:31			
MHS 73A (B.F.)				6:31			
MHS 95				6:47			
MHS 96				6:47			

Customer Remarks / Special Conditions / Possible Hazards:

Additional Instructions from Pace*:

Collected By: William A. Kotas (PSI)
Signature: [Signature]

Received by/Company (Signature): [Signature] PACE
Date/Time: 1/25/24 12:14

Received by/Company (Signature): [Signature] PACE
Date/Time: 1/25/24 12:14

Received by/Company (Signature): [Signature] PACE
Date/Time: 1/26 8:00

Received by/Company (Signature): [Signature] PACE
Date/Time: 1/26 8:00

Tracking Number: 1125/24 1714
 1125/24 8:20
 1126 4:10
 1126 8:00

Coolers: 3
Thermometer ID: TH1M
Correction Factor (°C): -0.4
Obs. Temp. (°C): 20.2
Corrected Temp. (°C): 20.6

Delivered by: In-Person Courier
 FedEx UPS Other

Page: 2 of 5

ENV-FRM-CORQ-0019_v01_082123 ©

Company Name: Intertek-PSI
Street Address: 17 British American Blvd, Latham NY 12110
Customer Project #: 8215304
Project Name: Mechanicville Central School District

Contact/Report To: William A. Kotas
Phone #: 518-377-9841
E-Mail: William.Kotas@intertek.com
Cc E-Mail:
Invoice To: PSI Latham NY Accounts Payable
Invoice E-Mail: LathamAR@intertek.com

Site Collection Info/Facility ID (as applicable):
Mechanicville Junior-Senior High School
Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET
Data Deliverables:

Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW
Rush (Pre-approval required): DW PWSID # or WW Permit # as applicable:
 ([] 12 Day [] 3 day [] 5 day [] Other)
Date Results Requested: Standard 10 business day
 Field Filtered (if applicable): [] Yes [] No

Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected (or Composite Start)		Res. CLZ	Number & Type of Containers	Sample Comment
			Date	Time			
MHS 97	DW	G		6:42		1	200.8 Drinking Water (Pb only)
MHS 98				6:42			
MHS 99				6:42			
MHS 101				6:50			
MHS 111A (B.F.)				6:50			
MHS 112				6:55			
MHS 117				6:56			
MHS 300C - F1				7:01			
MHS 300C - BF1							
MHS 300C - F2							

Customer Remarks / Special Conditions / Possible Hazards:

Additional Instructions from Pace*:

Coolers: 3-DX
Thermometer ID: TH211
Correction Factor (°C): 10.9
Obs. Temp. (°C): 20.2
Corrected Temp. (°C): 20.6

Tracking Number:

Date/Time: 1/25/24 12:14
Date/Time: 1/25 23:00
Date/Time: 1/26 4:10
Date/Time: 1/26 8:50

Delivered by: [] In-Person [X] Courier
 [] FedEx [] UPS [] Other

Page: 3 of 5



LAB USE ONLY - Affix Workorder/Login Label Here

Scan QR Code for Instructions

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Pace®
 Pace* Location Requested (City/State):
 Pace Analytical Long Island NY
 575 Broad Hollow Rd, Melville, NY 11747

Company Name: Intertek-PSI
Street Address: 17 British American Blvd, Latham NY 12110
Customer Project #: 8215304
Project Name: Mechanicville Central School District

Contact/Report To: William A. Kotas
Phone #: 518-377-9841
E-Mail: William.Kotas@intertek.com
Ct E-Mail:

Invoice To: PSI Latham NY Accounts Payable
Invoice E-Mail: LathamAR@intertek.com

Purchase Order # (if applicable): CR Borces
Quote #: CR Borces

Site Collection Info/Facility ID (as applicable):
 Mechanicville Junior-Senior High School

Time Zone Collected: [] AK [] JC [] MT [] CT [X] ET

Data Deliverables:
 [] Level II [] Level III [] Level IV
 [] EQUIS
 [] Other

County / State origin of sample(s): New York

Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW

Rush (Pre-approval required):
 DW PWSID # or WW Permit # as applicable:
 [] 2 Day [] 3 day [] 5 day [] Other

Date Results Requested: Standard 10 business day

Field Filtered (if applicable): [] Yes [] No

Analysis:

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected		Res. CL2	Number & Type of Containers	Sample Comment
			Date	Time			
MHS 301C-F1	DW	G	1/25/24	7:08		1	200.8 Drinking Water (Pb only)
MHS 301C-BF1							
MHS 301C-FZ							
MHS 301C-BFZ							
MHS 302-1							
MHS 302-2							
MHS 302-3							
MHS WRF							
MHS 162							
MHS 163							

Customer Remarks / Special Conditions / Possible Hazards:

Lead

Collected By: Printed Name: William A. Kotas (PSI)
 Signature: *William A. Kotas*

Requested by/Company (Signature): *William A. Kotas*

Received by/Company (Signature): *Maria Pace*

Received by/Company (Signature): *Maria Pace*

Received by/Company (Signature): *Maria Pace*

Received by/Company (Signature): *Maria Pace*

Additional Instructions from Pace*:

Coolers: Thermometer ID: Correction Factor (°C): Obs. Temp. (°C) Corrected Temp. (°C)
 3-22 TH11 40.0 70.2 70.6

Tracking Number: Date/Time: 1/25/24 12:14

Delivered by: [] In-Person [X] Courier
 [] FedEx [] UPS [] Other

Page: 4 of 5

Company Name: **Intertek-PSI**
 Street Address: **17 British American Blvd, Latham NY 12110**

Customer Project #: **8215304**
 Project Name: **Mechanicville Central School District**

Site Collection Info/Facility ID (as applicable):
Mechanicville Junior-Senior High School

Time Zone Collected: [] AK [] MT [] CT [X] ET
 Data Deliverables:

[X] Level II [] Level III [] Level IV
 [] EQUIS
 [] Other

* Matrix Codes (insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk (V), Other (OT)

Matrix *	Customer Sample ID	Comp / Grab	Collected (or Composite Start)		Composite End Date	Res. CL2	Number & Type of Containers	
			Date	Time			Plastic	Glass
DW	MHS 7 (FTN.)	G	1/25/24	7:25			1	
	MHS 2A (B.F.)			7:41				
	BG-52			7:42				
	BG-1							

200.8 Drinking Water (Pb only)

Proj. Mgr:	AccNum / Client ID:	Table #:	Profile / Template:	Prelog / Bottle Ord. ID:	Sample Comment
Loft Beyer					

Lab Use Only
 Preservation non-conformance identified for sample

CHAIN-OF-CUSTODY Analytical Request Document
 Chain-of-Custody is a LEGAL DOCUMENT - Complete all relevant fields

Contact/Report To: **William A. Kotas**
 Phone #: **518-377-9841**
 E-Mail: **William.Kotas@intertek.com**
 Cc E-Mail:

Invoice To: **PSI Latham NY Accounts Payable**
 Invoice E-Mail: **LathamAR@intertek.com**

Purchase Order # (if applicable):
CR Boces

Quote #: **CR Boces**

County / State origin of sample(s):
New York

Rush (Pre-approval required):
 [] 1-2 Day [] 3 day [] 5 day [] Other

Date Results Requested:
Standard 10 business day

Field Filtered (if applicable): [] Yes [] No
 Analysis:



Scan QR Code for instructions

Specify Container Size **

Identify Container Preservative Type ***

Analysis Requested

** Container Size: (1) 1L, (2) 500mL, (3) 250mL, (4) 125mL, (5) 100mL, (6) 40mL vial, (7) EnCare, (8) TerraCare, (9) Other

*** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other

Customer Remarks / Special Conditions / Possible Hazards:
 Lead

Collected By:
 Printed Name: **William A. Kotas (PSI)**
 Signature: *[Signature]*

Relinquished by/Company: (Signature) *[Signature]* **PSI**
 Date/Time: **1/25/24 12:14**
 Relinquished by/Company: (Signature) *[Signature]* **MPIN**
 Date/Time: **1/25/24 12:14**
 Relinquished by/Company: (Signature) *[Signature]* **MPIN**
 Date/Time: **1/26 8:50**
 Relinquished by/Company: (Signature) *[Signature]* **MPIN**
 Date/Time: **1/26 8:50**

Additional Instructions from Pace*:

Coolers: Thermometer ID: Correction Factor (°C): Obs. Temp. (°C) Corrected Temp. (°C)
 3-Box 44711 10.4 20.2 20.6

Tracking Number:

Date/Time: **1/25/24 12:14**

Date/Time: **1/25 23:00**

Date/Time: **1/26 4:10**

Date/Time: **1/26/24 8:50**

Delivered by: [] In-Person [X] Courier
 [] FedEx [] UPS [] Other

Page: **5** of **5**

Client: **Inter-Latham** Profile #: **10367**

Use Point Number Spreadsheet Multiday Project

Work ID: **Methuen Middle Central School District / 5** of **5** Add SCLOGFD to first sample for field charge

CCC Line Item	Matrix	Container Codes	Matrix
1			
2			
3			
4			
5			
6			
7			
8			
9			
10			
11			
12			
13			
14			
15			
16			
17			
18			
19			
20			
21			
22			
23			
24			
25			
26			
27			
28			
29			
30			
31			
32			
33			
34			
35			
36			
37			
38			
39			
40			
41			
42			
43			
44			
45			
46			
47			
48			
49			
50			
51			
52			
53			
54			
55			
56			
57			
58			
59			
60			
61			
62			
63			
64			
65			
66			
67			
68			
69			
70			
71			
72			
73			
74			
75			
76			
77			
78			
79			
80			
81			
82			
83			
84			
85			
86			
87			
88			
89			
90			
91			
92			
93			
94			
95			
96			
97			
98			
99			
100			

Code	Description	Matrix
BP1U	1L unreserved plastic	Water
BP3N	250mL HNO3 plastic	Solid
BP3C	250mL Sodium Hydroxide	Non-aqueous Liquid
AG2U	500mL unreserved amber glass	Oil
		Wipe
		Drinking Water

Code	Description	Matrix
BP1U	1L unreserved plastic	Water
BP3N	250mL HNO3 plastic	Solid
BP3C	250mL Sodium Hydroxide	Non-aqueous Liquid
AG2U	500mL unreserved amber glass	Oil
		Wipe
		Drinking Water

* Can also be a BPAN

Code	Description	Matrix
VG9T	40mL Na Thio amber vial	
DG9A	40mL Ascorbic acid/ maleic Acid vials	
DG9Y	Citrate/Na Thiosulfate 40mL	
DG5T	Na Thiosulfate 60mL vial	
DG9M	MonoChloric/Na Thio 60mL	
AG3U	250mL unreserved amber glass	
AG3T	Na Thiosulfate 250mL bottle	
BP1B	Na Thiosulfate Amber bottle	
AG1T	Na Thiosulfate 1L Amber	
AG1A	525-3 Chemical Blend	

Sender Initials _____

Additional Comments

WO#: 70285337
PM: LAB Due Date: **02/08/24**
CLIENT: INTER-LATHAM

WO#: 70285337

Client Name: INTER-LATHAM Project # _____

PM: LAB Due Date: 02/08/24
CLIENT: INTER-LATHAM

Courier: Fed Ex UPS USPS Client Commercial Pac Other

Tracking #: _____

Custody Seal on Cooler/Box Present: Yes No Seals intact: Yes No Temperature Blank Present: Yes No
 Packing Material: Bubble Wrap Bubble Bags Ziplo Non Other Type of Ice: Ice Blue None

Thermometer Used: JH211 Correction Factor: +0.4 Samples on ice, cooling process has begun
 Cooler Temperature (°C): 20.2 Cooler Temperature Corrected (°C): 20.6 Date/Time 5035A kits placed in freezer _____

Temp should be above freezing to 6.0°C

USDA Regulated Soil (N/A, water sample)

Did samples originate in a quarantine zone within the United States: AL, AR, CA, FL, GA, ID, LA, MS, NC, NM, NY, OK, OR, SC, TN, TX, or VA (check map)? Yes No

Did samples originate from a foreign source including Hawaii and Puerto Rico? Yes No

If Yes to either question, fill out a Regulated Soil Checklist (ENV-FRM-MELV-0076) and include with SCUR/COC paperwork.

Date and Initials of person examining contents: wli/26/24

	COMMENTS:
Chain of Custody Present: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	1.
Chain of Custody Filled Out: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	2.
Chain of Custody Relinquished: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	3.
Sampler Name & Signature on COC: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	4.
Samples Arrived within Hold Time: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	5.
Short Hold Time Analysis (<72hr): <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	6.
Rush Turn Around Time Requested <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No	7.
Sufficient Volume: (Triple volume provided for MS/MSD) <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	8.
Correct Containers Used: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	9.
-Pace Containers Used: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	
Containers Intact: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	10.
Filtered volume received for Dissolved tests <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	11. Note: if sediment is visible in the dissolved container.
Sample Labels match COC: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	12.
-Includes date/time/ID/Analysis Matrix: SL WT OIL OTHER	

Date and Initials of person checking preservation: wli/26/24

All containers needing preservation have been <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	13. <input type="checkbox"/> HNO ₃ <input type="checkbox"/> H ₂ SO ₄ <input type="checkbox"/> NaOH <input type="checkbox"/> HCl
pH paper Lot # <u>22782201</u>	Sample #
All containers needing preservation are found to be in compliance with method recommendation? (HNO ₃ , H ₂ SO ₄ , HCl, NaOH > 9 Sulfide, <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A NAOH > 12 Cyanide)	
Exceptions: VOA, Coliform, TOC/DOC, Oil and Grease, DRO/8015 (water).	Initial when completed: _____ Lot # of added preservative: _____ Date/Time preservative added: _____
Per Method, VOA pH is checked after analysis	
Samples checked for dechlorination: <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	14.
KI starch test strips Lot #	Positive for Res. Chlorine? Y N
Residual chlorine strips Lot #	
SM 4500 CN samples checked for sul <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	15.
Lead Acetate Strips Lot #	Positive for Sulfide? Y N
Headspace in VOA Vials (>6mm): <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	16.
Trip Blank Present: <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	17.
Trip Blank Custody Seals Present <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	

DATE AND INITIALS OF PERSON COMPLETING SECOND REVIEW: _____

Client Notification/ Resolution: _____

Field Data Required? Y / N

Person Contacted: _____

Date/Time: _____

Comments/ Resolution: _____

* PM (Project Manager) review is documented electronically in LIMS.